



COLLEGE of AMERICAN  
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

ORIGINAL EVALUATION

**ID3-B 2025 Nucleic Acid Amp, Respiratory Ltd (ID3)**

|                         |  |                         |                                  |
|-------------------------|--|-------------------------|----------------------------------|
| <b>INSTITUTION:</b>     | Credo Diagnostics Biomedical Pte Ltd<br>Taipei TW 10572        |                         |                                  |
| <b>ATTENTION:</b>       | Chia Hsin Su MD  |                         |                                  |
| <b>CAP NUMBER:</b>      | 9867369-01   | <b>Kit Sequence:</b>    | 1                                |
| <b>KIT INFORMATION:</b> | Kit #:<br>39181771   | Kit Mailed:<br>7/7/2025 | Original Evaluation:<br>8/6/2025 |
| <b>COPIED TO:</b>       | CAP  |                         |                                  |
| <b>LEGEND:</b>          | Exception Reason Codes appearing in this evaluation:<br><NONE> |                         |                                  |

Reviewed By

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**CAP Number:** 9867369-01      **Kit Sequence 1**  
**Institution:** Credo Diagnostics Biomedical Pte Ltd  
**Attention:** Chia Hsin Su MD  
**City / State:** Taipei TW 10572

**Kit #:** 39181771  
**Kit Mailed:** 7/7/2025  
**Original Evaluation:** 8/6/2025

**EVALUATION**  
**ORIGINAL**

**ID3-B 2025 Nucleic Acid Amp, Respiratory Ltd (ID3)**

| <b>Test Method</b>                | <b>Specimen</b> | <b>Your Result</b>      | <b>Intended Response</b> | <b>Your Grade</b> |
|-----------------------------------|-----------------|-------------------------|--------------------------|-------------------|
| <b>Influenza A virus</b><br>OTHER | <b>ID3-06</b>   | DETECTED (POSITIVE)     | DETECTED (POSITIVE)      | Good              |
|                                   | <b>ID3-07</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |
|                                   | <b>ID3-08</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |
|                                   | <b>ID3-09</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |
|                                   | <b>ID3-10</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |
| <b>Influenza B virus</b><br>OTHER | <b>ID3-06</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |
|                                   | <b>ID3-07</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |
|                                   | <b>ID3-08</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |
|                                   | <b>ID3-09</b>   | DETECTED (POSITIVE)     | DETECTED (POSITIVE)      | Good              |
|                                   | <b>ID3-10</b>   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)  | Good              |

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



CAP Number: 9867369-01

Kit Sequence 1

Kit #: 39181771

Institution: Credo Diagnostics Biomedical Pte Ltd

Kit Mailed: 7/7/2025

Attention: Chia Hsin Su MD

Original Evaluation: 8/6/2025

City / State: Taipei TW 10572

EVALUATION  
ORIGINAL

## ID3-B 2025 Nucleic Acid Amp, Respiratory Ltd (ID3)

| Test Method         | Specimen | Your Result             | Intended Response                           | Your Grade |
|---------------------|----------|-------------------------|---|------------|
| RSV<br>OTHER        | ID3-06   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |
|                     | ID3-07   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |
|                     | ID3-08   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |
|                     | ID3-09   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |
|                     | ID3-10   | DETECTED (POSITIVE)     | DETECTED (POSITIVE)                         | Good       |
| SARS-CoV-2<br>OTHER | ID3-06   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |
|                     | ID3-07   | DETECTED (POSITIVE)     | DETECTED (POSITIVE)<br>PRESUMPTIVE POSITIVE | Good       |
|                     | ID3-08   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |
|                     | ID3-09   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |
|                     | ID3-10   | NOT DETECTED (NEGATIVE) | NOT DETECTED (NEGATIVE)                     | Good       |

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**CAP Number:** 9867369-01      **Kit Sequence 1**  
**Institution:** Credo Diagnostics Biomedical Pte Ltd  
**Attention:** Chia Hsin Su MD  
**City / State:** Taipei TW 10572

**Kit #:** 39181771  
**Kit Mailed:** 7/7/2025  
**Original Evaluation:** 8/6/2025

**EVALUATION**  
**ORIGINAL**

**ID3-B 2025 Nucleic Acid Amp, Respiratory Ltd (ID3)**

**CAP**

| CAP #:                         |  |         |     | Subspecialty : Virology |            |         |                   |     |            |         |     |     |
|--------------------------------|--|---------|-----|-------------------------|------------|---------|-------------------|-----|------------|---------|-----|-----|
| Regulated Analyte              | Proficiency Event  |         |     | Proficiency Event       |            |         | Proficiency Event |     |            |         |     |     |
|                                | Test Event   | Score   | %   | Wgt                     | Test Event | Score   | %                 | Wgt | Test Event | Score   | %   | Wgt |
| Viral Identification           | ID3-C  | 20 / 20 | 100 | 20                      | ID3-A      | 20 / 20 | 100               | 20  | ID3-B      | 20 / 20 | 100 | 20  |
| Performance By Mailing         | Satisfactory   |         |     | Satisfactory            |            |         | Satisfactory      |     |            | 100     |     |     |
| Regulatory Status for Virology | Cumulative Regulatory Performance Interpretation: Successful |         |     |                         |            |         |                   |     |            |         |     |     |

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.